

Certificate of Analysis

Product name: SARS-CoV-2 RT-PCR Multiplex Diagnostic Kit **Manufacture date:** 22 Jan 2021
Product code: NAT-01 **Expiry date:** 01-22
Batch number: NAT-01-210102 **QC release date:** 25 Jan 2021

Test	Specification	Result Yes / No
Appearance:		
NAR-01 Master Mix	Blue liquid, in 1.5mL vial, blue label	Yes
NAR-02 Enzyme Mix	Clear liquid, in 0.5mL vial, red label	Yes
NAR-03 Multiplex Mix	Clear liquid, in 0.5mL vial, yellow label	Yes
NAR-04 Positive Control	Clear liquid, in 0.5mL vial, green label	Yes
NAR-05 Negative Control	Clear liquid, in 1.5mL vial, white label	Yes
CT (Cq) value:		
N gene	≤ 38	Yes
E gene	≤ 38	Yes
ORF1ab gene	≤ 38	Yes
Rnase P gene	≤ 38	Yes
NTC	No Ct (excluded) or >38	Yes

This batch has been manufactured and tested according to the applicable manufacturing and quality control documents and has met all product release requirements.



Production Manager
Jipsa Mary James