

# Certificate of Analysis

**Product name:** SARS-CoV-2 RT-PCR Multiplex Diagnostic Kit      **Manufacture date:** 24 Sep 2021  
**Product code:** NAT-01      **Expiry date:** 11-22  
**Batch number:** NAT-01-210904      **QC release date:** 28 Sep 2021

Test	Specification	Result Yes / No
<b>Appearance:</b>		
NAR-01 Master Mix	Blue liquid, in 1.5mL vial, blue label	Yes
NAR-02 Enzyme Mix	Clear liquid, in 0.5mL vial, red label	Yes
NAR-03 Multiplex Mix	Clear liquid, in 0.5mL vial, yellow label	Yes
NAR-04 Positive Control	Clear liquid, in 0.5mL vial, green label	Yes
NAR-05 Negative Control	Clear liquid, in 1.5mL vial, white label	Yes
<b>CT (Cq) value:</b>		
N gene	≤ 38	Yes
E gene	≤ 38	Yes
ORF1ab gene	≤ 38	Yes
Rnase P gene	≤ 38	Yes
NTC	No Ct (excluded) or >38	Yes

This batch has been manufactured and tested according to the applicable manufacturing and quality control documents and has met all product release requirements.



**Production Manager**  
**Jipsa James**